

MW570

LISTERIA ISOLATION TRANSWAB®

Listeria Isolation Transwab is designed to be used alongside traditional selective methods to improve the quality system and minimise the risk of Listeria contamination. This simple to use diagnostic test can be applied anywhere in the environment where the presence of *Listeria spp.* would be critical. *Listeria spp.* and specifically *Listeria monocytogenes* are rapidly becoming the most important pathogen in the Food Industry; regulatory bodies from around the world are insisting that all food products are Listeria free.

Listeria Isolation Transwab works on an enhanced esculin media formulation. The hydrolysis of esculin gives a distinctive black/brown precipitate. Inhibitors are present in the media which will inhibit the growth of non *Listeria spp.* Only *Listeria spp.* will grow within the media to produce a black/brown precipitate due to esculin hydrolysis (table 3)⁵.

If any black media tubes are found as soon as the box is opened DO NOT USE.

Instructions for Use

1. Peel back wrapper to expose both caps.
2. Remove swab and sample test site by rubbing swab over as large an area as practicable.
3. Remove cap of culture tube with thumb and forefinger and discard.
4. Insert swab in culture tube and push down fully to immerse the swab completely.
5. Fill in time, date and site details.
6. Incubate at 37°C for up to 48 hours.
7. Read results. A positive result is indicated by a colour change from Light Brown Agar to Black/Dark Brown commencing around the bud, any colour change is significant. Negative result shows no colour change.
8. Record results and dispose of tube.

Shelf Life/Expiry Date

The expiry date is 12 months from the month of manufacture and is printed onto the Tube label, Peel pouch and outer Boxes.

Storage

Listeria Isolation Transwab® should be stored in a dry place between +5°C to +25°C. **Do NOT freeze.**

Sterilisation Method

Listeria Isolation Transwab® has been sterilised by irradiation.

References

1. Billie, J. and Doyle, M.P. *Listeria and Erysipelothrix*. In: *Manual of Clinical Microbiology*, edited by Balows, A., Hausler Jr., W.J. Hermann, K.L., Isenberg, H.D., and Shadomy, H.J., Washington, DC 2000: American Society for Microbiology, 1992 p. 287-295.
2. Newton, L., Hall, S.M., Pelerin, M., and Mclauchlin, J. Listeriosis surveillance: 1991. *CDR* 12:R142-R144, 1992.
3. Newton, L., Hall, S.M., and Mclauchlin, J. Listeriosis surveillance: 1992. *CDR* 3:10:R144-R146, 1993.
4. Sprenger, R.A. *Hygiene for Management*, Doncaster, England:Highfield Publications, 1991 pp. 3-303.
5. Turner, R. A Study of MW570 Listeria Transwab, Personal Communication: 1994 Abtek Biologicals Ltd, Liverpool U.K.

Precautions When Swabbing!

1. When sampling an area always use the same swabbing technique.
2. Always try to sample the same surface area (100cm² or 200cm²), and be consistent in your sampling.
3. The moisture level on the surface should not vary at different sample times. It may be preferable with dry surfaces to wet the swab in a suitable wetting agent such as Ringers, Tween Lecithin Thiosulphate Ringers or Tween Peptone water.

Medical Wire & Equipment

Corsham, Wiltshire, SN13 9RT

Tel: +44 (0) 1225 810361

Fax: +44 (0) 1225 810153


info@mwe.co.uk

www.mwe.co.uk

Battery of Organisms Tested against MW570 Listeria Isolation Transwab®

<u>Organism</u>	<u>C.F.U. in Sample</u>	<u>Colour Change</u>
<i>Listeria monocytogenes</i> NCTC 11994	7	Black
<i>Listeria monocytogenes</i> NCTC 5214	19	Black
<i>Listeria monocytogenes</i> NCTC 7973	15	Black
<i>Listeria ivanovii</i>	10	Black
<i>Listeria innocua</i> NCTC 11288	9	Black
<i>Listeria seeligeri</i>	1 x 10 ⁴	Black
<i>Listeria welshimeri</i> NCTC 11857	1 x 10 ³	Black
<i>Listeria murrayi</i>	1 x 10 ⁵	Black
<i>Listeria grayi</i>	1 x 10 ⁵	Black
<i>Staphylococcus aureus</i> NCTC 6571	4 x 10 ⁸	No Change
<i>Strep faecalis</i> NCTC 775	1.5 x 10 ⁵	No Change
<i>Aeromonas hydrophilla</i> NCTC 1767	1.5 x 10 ⁷	No Change
<i>Bacillus subtilis</i> NCTC 10400	4 x 10 ⁷	No Change
<i>Bacilus pumilis</i> NCTC 10327	1 x 10 ⁷	No change
<i>Bacilus cereus</i> NCTC 10320	2 x 10 ⁶	No change
<i>Escherichia coli</i> NCTC 9001	1 x 10 ⁷	No change
<i>Escherichia coli</i> NCTC 10418	6 x 10 ⁷	No change
<i>Klebsiella pneumoniae</i>	3 x 10 ⁷	No change
<i>Klebsiella aerogenes</i> NCTC 7418	1 x 10 ⁸	No change
<i>Klebsiella aerogenes</i> NCTC 11228	3 x 10 ⁷	No change
<i>Proteus vulgaris</i> NCTC 1683	1 x 10 ⁸	No change
<i>Proteus mirabilis</i> NCTC 841	5 x 10 ⁸	No change
<i>Citrobacter freundii</i>	1.2 x 10 ⁸	No change
<i>Citrobacter diversus</i>	3 x 10 ⁶	No change
<i>Enterobacter agglomerans</i>	1.1 x 10 ⁸	No Change
<i>Morganella morgani</i>	1.4 x 10 ⁸	No Change
<i>Serratia liquefaciens</i>	1 x 10 ⁸	No Change
<i>Serratia marcescens</i>	1 x 10 ⁷	No Change
<i>Yersinia enterocolitica</i>	1 x 10 ⁷	No Change
<i>Enterobacter cloacae</i>	3 x 10 ⁷	No Change
<i>Salmonella typhimurium</i> NCTC 74	3 x 10 ⁸	No Change
<i>Shigella flexneri</i>	8 x 10 ⁶	No Change
<i>Pseudomonas aeruginosa</i>	9 x 10 ⁷	No Change
<i>Pseudomonas putida</i> NCTC 10936	1.5 x 10 ⁷	No Change
<i>Pseudomonas fluorescens</i> NCTC 10038	1 x 10 ⁷	No Change
<i>Pseudomonas maltophilia</i> NCTC 102157	6 x 10 ⁷	No Change
<i>Pseudomonas vesicularis</i> NCTC 10900	4 x 10 ⁶	No Change
<i>Pseudomonas cepacia</i> NCTC 10743	3.5 x 10 ⁷	No Change
<i>Pseudomonas putrefaciens</i> NCTC 10735	3 x 10 ⁶	No Change
<i>Candida albicans</i>	2 x 10 ⁶	No Change

Example of surface area to be swabbed



10 cm